



# Ref. Certif. No.

IL1790

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME	
CB TEST CERTIFICATE	
Product	Capacitor charging module
Name and address of the applicant	OEM Tech Sp.z o.o. 58 Marszalkowska st., 00-545, Warsaw, Poland
Name and address of the manufacturer	OEM Tech Sp.z o.o. 58 Marszalkowska st., 00-545, Warsaw, Poland
Name and address of the factory	OEM Tech Sp.z o.o. ul. 42 Pułku Piechoty 8, 15-181, Białystok, Poland
Note: When more than one factory, please report on page 2	Additional Information on page 2
Ratings and principal characteristics	Input: 200-240V~, 50-60Hz, 24.0-19.5A Output 800Vd.c.
Trademark (if any)	
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	PCA-40-800V-PD
Additional information (if necessary may also be reported on page 2)	Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012/AMD2:2020, including National Differences: CA, JP, US
As shown in the Test Report Ref. No. which forms part of this Certificate	7412311725
This CB Test Certificate is issued by the National Certification Body	
The Standards Institution of Isl 42 Chaim Levanon St., Tel Aviv 6997701, Israel	rael
Date: March 30, 2025	Signature: Moria Stern H.S.



Date: 2021-03-24

**UA-0187** 

# **CB TEST CERTIFICATE** Product Capacitor charging power supply OOO "OEM Tech", Name and address of the applicant Odoevskogo Str.129-106, 220018 Minsk, Belarus 000 "OEM Tech". Name and address of the manufacturer Odoevskogo Str.129-106, 220018 Minsk, Belarus 000 "OEM Tech". Name and address of the factory Odoevskogo Str. 129-106, 220018 Minsk, Belarus Additional Information on page 2 Note: When more than one factory, please report on page 2 Ratings and principal characteristics Input:200-240VAC, 50-60Hz, 24-19,5 A, Class I Output: 1000VDC Trademark / Brand (if any) **OEM Tech** Customer's Testing Facility (CTF) Stage used Model / Type Ref. PCA-40-1000V-PD Additional information (if necessary may also be The risk management requirements of the standard reported on page 2) were not addressed Additional Information on page 2 A sample of the product was tested and found to be in conformity with IEC 60601-1:2005/AMD1:2012 inclusive National differences for USA, Canada As shown in the Test Report Ref. No. which 1322-2-2020 forms part of this Certificate This CB Test Certificate is issued by the National Certification Body

4, Metrologichna, str., 03143 Kyiv, Ukraine Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical Equipment UkrTEST of SE "UKRMETRTESTSTANDART"

**Victor Rymer** 



**UA-0192** 

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (EQEE) OB SCHEME

#### **CB TEST CERTIFICATE**

Product

Name and address of the applicant

Name and address of the manufacturer

Name and address of the factory

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

Trademark / Brand (if any)

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

As shown in the Test Report Ref. No. which forms part of this Certificate

## **Capacitor charging power supply**

OOO "OEM Tech", Odoevskogo Str.129-106, 220018 Minsk, Belarus OOO "OEM Tech", Odoevskogo Str.129-106, 220018 Minsk, Belarus OOO "OEM Tech", Odoevskogo Str.129-106, 220018 Minsk, Belarus

Input:200-240VAC, 50-60Hz, 24-19.5 A Output: 1000V

**OEM Tech** 

### PCA-40-1000V-PD

The risk management requirements of the standard were not addressed

IEC 60601-1-2:2014 inclusive National differences for Canada

2018-5-2020

This CB Test Certificate is issued by the National Certification Body

4, Metrologichna, str., 03143 Kyiv, Ukraine Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical Equipment UkrTEST of SE "UKRMETRTESTSTANDART"

Date: 2021-06-09

Signature:

Victor Rymer



UA-0185

# IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

#### **CB TEST CERTIFICATE**

Product

Name and address of the applicant

Name and address of the manufacturer

Name and address of the factory

Note: When more than one factory, please report on page-2

Ratings and principal characteristics

Trademark / Brand (if any)

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

As shown in the Test Report Ref. No. which forms part of this Certificate

Capacitor charging power supply

OOO "OEM Tech", Odoevskogo Str.129-106, 220018 Minsk, Belarus OOO "OEM Tech", Odoevskogo Str.129-106, 220018 Minsk, Belarus

OOO "OEM Tech", Odoevskogo Str.129-106, 220018 Minsk, Belarus

Additional Information on page 2

Input:200-240VAC, 50-60Hz, 21-18 A, Class I Output: 1300VDC

**OEM Tech** 

PCA-40-1300V-CD

Additional Information on page 2

IEC 60601-1:2005/AMD1:2012 inclusive National differences for USA, Canada

2019-2-2020

This CB Test Certificate is issued by the National Certification Body

4, Metrologichna, str., 03143 Kyiv, Ukraine Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical Equipment UkrTEST of SE "UKRMETRTESTSTANDART"

Date: 2021-02-24

**Victor Rymer** 

1 4



OEM Tech LLC. declares under its sole responsibility that the product(s) to which this declaration relates, is in conformity with the following directives, standard(s) and other normative document(s).

List of models:	PCA-40-1000V-PD
Directives:	Medical Device Directive 93/42/EEC Electromagnetic Compatibility Directive 2014/30/EU Restriction of Hazardous Substances Directive 2011/65/EU
Product Safety Standards:	IEC 60601-1:2005/AMD1:2012
EMC Standards:	IEC 60601-1-2:2014

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

June 01, 2021

Dmitry Kamlyuk, CEO

F

OEM Tech LLC 129 Odoevskogo str. 220018 Minsk Belarus T. +375 17 3224054 F. +375 17 3224064 info@oem-tech.by www.oemtex.com